



Phoenix Group®

Vía España, Plaza Dorchester #307/309, Ciudad de Panamá
Manzanillo International Terminal, Colon, Panamá

For sales and pricing information contact:

Michael McNicholas, Director
mmcnicholas@phoenixgrouppanama.com



AAMI 3 ●●●●●
Moderate degree of invasion

— Plus Medical Gown L —

CVE: CLIN 0007



Does not contain latex




Do not expose to high temperatures



Flammable

CONTENT




Medical Gown,
Size Large

CAUTION: The contents of this package, like that of any fabric, may burn if exposed to a flame or extremely hot objects. Do not use in the presence of flammable anesthesia. The materials in this package are found in the Flammability Standard for Wearable Textiles CPSC 16 CFR part. 1640.
Waterproof reinforcements have been tested and found to meet the following standards: ASTM F1670 (Standard Test Method for Strength of Materials Used in Protective Clothing from Penetration by Synthetic Blood) and ASTM 1671 (Protection from Penetration Protection by Blood with Pathogens (Using Phi-X174 Bacteriological Penetration as test system).

The measurement of some of the garments in this package may vary more / less 5 cm.



Click here to see the instructions for the correct use of this product.





Phoenix Group®

Vía España, Plaza Dorchester #307/309, Ciudad de Panamá
Manzanillo International Terminal, Colon, Panamá

Excelencia en Protección Quirúrgica

TECHNICAL DATA SHEET

GENERAL SPECIFICATIONS

Plus Medical Gown

DESCRIPTION

Medical gown package for single-use. Non- alcohol repellent. **Certified** under international and national standards such as ISO 13485, ISO 9001 and NOM 241. Elaborated and manufactured according to NE 13795, AAMI, ASTM and EDANA standards <https://www.edana.org/discover-nonwovens/products-applications/medical>. All the materials used in this product are specific for medical surgical use, hypoallergenic, latex-free, **antistatic**, **anti-reflective** and resistant without **particle** detachment. The fabrics used exceed the repellency / absorbency, impermeability and breathability parameters established by standards EN-20811, 22610, ISO 11737-1, 9073-10, EN-13938-1, EN 29073-3, NMX-A-303- INNTEX-2000, ASTM F1670 and ASTM F1671.

QUANTITY	COMPONENT	MEASUREMENTS			TOL	UNITS	RAW MATERIAL	
		CLIN 007 L	CLIN 008 XL	CLIN 009 XXL				
A	1	Medical gown	135 x 150	140 x 165	150 x 175	+/-2	cm2	
		Body medical gown	135 x 150	140 x 165	150 x 175	+/-2	cm2	44 gsm;100% Polypropylene SMMS Blue AAMI 3
		Medical gown sleeve	65 x 60	65 x 60	75 x 68	+/-2	cm2	44 gsm;100% Polypropylene SMMS Blue AAMI 3
		Short lace medical gown	60 x 04	60 x 04	60 x 04	+/-2	cm2	44 gsm;100% Polypropylene SMMS Blue AAMI 3
		Large lace medical gown	70 x 04	70 x 04	80 x 04	+/-2	cm2	44 gsm;100% Polypropylene SMMS Blue AAMI 3
		Cuff	17 X 17	17 X 17	17 X 17	+/-0.5	cm2	White Cardigan
		Collar medical gown	78	78	78	+/-2	cm1	White bias bindings

Characteristics	Units	MAX	Hindernis	MIN
Grab Machine Resistance	(lb/in)	28	26	22
Machine Elongation	(%)	58	54.2	41
Transv resistance Grab	(lb/in)	27	21.1	17.3
Trans Elongation	(%)	62	59.9	37
Air Perm	m3/m2/min	25	22	8
HSB (water column)	mbar	99	95.7	70
Alcohol repellency	N/A	N/A	N/A	N/A
Abrasion	mg	1.00	0.05	0.02
Electrostatics	s	3.00	0.03	0.00

PACKING

1	XCH Transparent Packaging Bag 29 x 20 cm	-	pz
1	Medical Gown Plus Hindernis Cover 105 x 148.5mm	-	pz



SPECIFICATIONS

Product after use treat as HBIW (handling of hazardous biological infectious waste).

Specifications indicated in the Technical Data Sheets and Hindernis Certificates

I.Q. SERGIO GUARNERO MENENDEZ
HEALTH MANAGER
CED.NO. 3609842



Product not currently approved by the FDA.