



Phoenix Group®

Vía España, Plaza Dorchester #307/309, Ciudad de Panamá
Manzanillo International Terminal, Colon, Panamá

For sales and pricing information contact:

Michael McNicholas, Director
mmcnicholas@phoenixgrouppanama.com



AAMI 3 
Moderate degree of invasion

— Plus Medical Gown L —
Sterile

CVE: CLIN 0001



Does not
contain latex




Do not expose to
high temperatures



Flammable

CONTENT



Medical Gown,
Size Large

CAUTION: The contents of this package, like that of any fabric, may burn if exposed to a flame or extremely hot objects. Do not use in the presence of flammable anesthetics. The materials in this package are found in the Flammability Standard for Wearable Textiles CPSC 16 CFR part. 1640. Waterproof reinforcements have been tested and found to meet the following standards: ASTM F 1670 (Standard Test Method for Strength of Materials Used in Protective Clothing from Penetration by Synthetic Blood) and ASTM 1671 (Protection from Penetration Protection by Blood with Pathogens (Using Phi X 174 Bacteriological Penetration as test system).

The measurement of some of the garments in this package may vary more / less 5 cm.



Click here to see the
instructions for the
correct use of this
product.





Phoenix Group®

Vía España, Plaza Dorchester #307/309, Cuidad de Panamá
Manzanillo International Terminal, Colon, Panamá

Excelencia en Protección Quirúrgica

TECHNICAL DATA SHEET

GENERAL SPECIFICATIONS

Plus Medical Gown STERILE

DESCRIPTION

STERILE Medical gown package for single-use. **Certified** under international and national standards such as ISO 13485, ISO 9001 and NOM 241. Elaborated and manufactured according to NE 13795, AAMI, ASTM and EDANA standards <https://www.edana.org/discover-nonwovens/products-applications/medical>. All the materials used in this product are specific for medical surgical use, hypoallergenic, latex-free, **antistatic**, **anti-reflective** and resistant without **particle** detachment. The fabrics used exceed the repellency / absorbency, impermeability and breathability parameters established by standards EN-20811, 22610, ISO 11737-1, 9073-10, EN-13938-1, EN 29073-3, NMX-A-303- INNTEX-2000, ASTM F1670 and ASTM F1671.

Gamma radiation sterilized product.

QUANTITY	COMPONENT	MEASUREMENTS			TOL	UNITS	RAW MATERIAL
		CLIN 001 L	CLIN 002 XL	CLIN 003 XXL			
A	1	Medical gown	135 x 150	140 x 165	150 x 175	+/-2	cm2
		Body medical gown	135 x 150	140 x 165	150 x 175	+/-2	cm2
		Medical gown sleeve	65 x 60	65 x 60	75 x 68	+/-2	cm2
		Short lace medical gown	60 x 04	60 x 04	60 x 04	+/-2	cm2
		Large lace medical gown	70 x 04	70 x 04	80 x 04	+/-2	cm2
		Cuff	17 x 17	17 x 17	17 x 17	+/-0.5	cm2
		Collar medical gown	78	78	78	+/-2	cm1
							44 gsm;100% Polypropylene SMMS Blue AAMI Lvl 3 Liquid Barrier Protection
							44 gsm;100% Polypropylene SMMS Blue AAMI Lvl 3 Liquid Barrier Protection
							44 gsm;100% Polypropylene SMMS Blue AAMI Lvl 3 Liquid Barrier Protection
							44 gsm;100% Polypropylene SMMS Blue AAMI Lvl 3 Liquid Barrier Protection
							White Cardigan
							White bias bindings

Characteristics	Units	MAX	Hindernis	MIN
Grab Machine Resistance	(lb/in)	28	26	22
Machine Elongation	(%)	58	54.2	41
Transv resistance Grab	(lb/in)	27	21.1	17.3
Trans Elongation	(%)	62	59.9	37
Air Perm	m3/m2/min	25	22	8
HSB (water column)	mbar	99	95.7	70
Alcohol repellency	Kit #	9	7	3
Abrasion	mg	1.00	0.05	0.02
Electrostatics	s	3.00	0.03	0.00

PACKING

1	XCH Transparent Packaging Bag 29 x 20 cm	-	pz
1	Medical Gown Plus Hindernis Cover 105 x 148.5mm	-	pz



SPECIFICATIONS

Product after use treat as HBIW (handling of hazardous biological infectious waste).



Product not currently approved by the FDA.

I.Q. SERGIO GUARNERO MENENDEZ
HEALTH MANAGER
CED.NO. 3609842